

	Clinical Pathology Accreditation (UK) Ltd	Requirement for Exclusion Zones Around Centrifuges
---	--	---

BSEN61010-2-020 Safety Requirements for Electrical Equipment for measurement, Control and Laboratory use – Part 2-020 Particular Requirements for Laboratory Centrifuges.

The document states that an exclusion zone is needed if there is a requirement within the manufacturers' instructions.